Technical Information Bulletin

Subject: Duo Processing Mode **BULLETIN NO.: 136**

[ ] Factory Modification [ ] Field Modification [ ] Safety Information [X] Technical Information [ ] General Information [ ] Adv. Tech Information	Applies to: GR1750	Modification Schedule: [ ] TBS [ ] Next Call [ ] Next PM [X] As Necessary	Factory Installed_____ Field Installed_____ FittingTime _____

Background

If the Duo Processing Mode is reguired on the GR1750, it must be enabled.

Areas of Reference

- Dip Switch 3 on the System PCB, must be set correctly to enable the Duo Processing Mode.

Solution

- Locate the System PCB 2 of the GR1750.
- Locate SW3 in the upper left corner of System PCB 2.
- Set DIP# 3 of SW3 to “ON” to enable the Duo Processing Mode.

Note: These steps will be required when replacing the System PCB 2 on a GR1750 that has the Duo Processing Mode selected.



System PCB 2



Issue Date:	Supersedes	Dated:	Page No.
03/21/96			1 of 1